



NACI Certificate



Certificate NO. 197

شرکت مهندسی آسا آزمون سنجش ابزار

CERTIFICATE OF CALIBRATION



Location

Issued by: AZMOON SANJESH ABZAR

Date of issue: 1400-11-02

Certificate NO: 001028096

NO32-POONAK BLVD-ASHRAFI ESFAHANI ST-TEHRAN-IRAN

Telefax: +9821 44445009 / +9821 44401554

Email: asacalibration@yahoo.com

http://www.asa-calibration.com

page 1 of 1 pages

Calibration date:

1400-11-02

Next calibration:
(Customer Request)

1402-11-02

APPLICANT: بی جی پارس

INSTRUMENT DESCRIPTION:

Instrument: Tape measure (متر فلزی)

Manufacturer: ASSIST

Serial NO: ---

Identification NO: ---

Measuring range: 0 to 3.5 M

Resolution: 1 mm

BASIS OF CALIBRATION:

1-Iranian Defense Standard IDS-GDL-0368

2-Calibration Instruction: LDW-16

3-International Standard: JIS B 7512

ENVIRONMENT:

■ Temperature: 20±2 °C

■ Humidity: 35% ~ 55%.

REFERENCE EQUIPMENTS:

1-Rule Calibrator / Serial NO: 65051CA1110002143

2- Surface plate grade 0 / Serial NO: JJG117-91

TRACEABILITY:

TRACEABIL to ML-10 Laser / SN: G92339,UKAS

Nominal value of the length مقدار اسمی	Actual value مقدار طول واقعی	Deviation انحراف از مقدار اسمی	Permissible deviation حداکثر خطای مجاز	EXP Uncertainty عدم قطعیت
millimeter	millimeter	millimeter	± millimeter	± millimeter
500	500.1	0.1	0.4	0.1
1000	1000.2	0.2	0.5	
1500	1500.1	0.1	0.5	
2000	2000.3	0.3	0.6	
2500	2500.1	0.1	0.7	
3000	3000.1	0.1	0.8	

CALIBRATION TECHNICIAN

VALID

NACI/Lab/197

APPROVED SIGNATORY

Technical Manager

LABORATORY STAMP



M.Zare

ID NO. LOS-F358-00