



NACI Certificate



شرکت مهندسی آسا آزمون سنجش ابزار

CERTIFICATE OF CALIBRATION



Location

Issued by: **AZMOON SANJESH ABZAR**

Date of issue: 1400-11-02

Certificate NO: **001028094**

NO32-POONAK BLVD-ASHRAFI ESFAHANI ST-TEHRAN-IRAN

Telefax: +9821 44445009 / +9821 44401554

Email: asacalibration@yahoo.com

<http://www.asa-calibration.com>

page 1 of 2 pages

Calibration date:

1400-11-02

Next calibration:
(Customer Request)

1402-11-02

APPLICANT: بی جی پارس

INSTRUMENT DESCRIPTION: Instrument: (1set of 5) Calibration foil

Manufacturer: -----

Serial NO: ---

IDentification NO: ---

BASIS OF CALIBRATION: 1-American society for testing and materials ASTM B499
2-Calibration Instruction: LDW-20

CALIBRATION METHOD: The central thickness of each calibration foil is determined, using a displacement measurement transducer. For determining the variation of thickness distortion is measured at four corner points about 5 mm from the face edges.

ENVIRONMENT: ■ Temperature: 20±2 °C
■ Humidity: 35% ~ 55%.

REFERENCE EQUIPMENTS: 1-Universal length measuring machine/Serial NO:101015

TRACEABILITY: Traceable to JCSS 0022 by Certificate NO:42992

QUALITY SYSTEM: Based on ISO / IEC 17025:2005

CALIBRATION TECHNICIAN



APPROVED SIGNATORY

Technical Manager

LABORATORY STAMP



M.Zare

ID NO. LOS-F362-00